

## TR-FRET ERR gamma Coactivator Assay Kit, rabbit

### Product Information

---

**Cat**

Kit-0911

---

**Common Name**

TR-FRET

---

**Cat.No.**

Kit-0911

---

**Description**

The TR-FRET Estrogen Related Receptor (ERR) gamma Coactivator Assay Kit provides a sensitive and robust method for high-throughput screening of potential ERR gamma ligands as antagonists/inverse agonists of coactivator displacement. The homogeneous mix-and-read assay format results in a ligand EC<sub>50</sub> composite value representing the amount of ligand required to bind to receptor, effect a conformational change, and displace coactivator peptide.

#### How it works

The TR-FRET ERR gamma Coactivator Assay Kit includes a terbium (Tb)-labeled anti-GST antibody, a fluorescein-labeled coactivator peptide, and a human ERR gamma ligand-binding domain (ERR gamma-LBD) that is tagged with glutathione-S-transferase (GST). To assay, ERR gamma-LBD is added to test compounds, followed by addition of a mixture of the fluorescein-coactivator peptide and Tb-anti- GST antibody. After room temperature incubation, the TR-FRET 520: 495 emission ratio is calculated and used to determine the EC<sub>50</sub> from a dose response curve of the compound.

---

**Applications**

Nuclear Receptor Assay

---

**Storage**

Tb-anti-GST Antibody (Rabbit): Store at -20 °C

ERR gamma-LBD, GST: Store at -80 °C

Fluorescein Coactivator Peptide: Store at -20 °C

TR-FRET Coregulator Buffer B: Store at 4 °C



## TR-FRET ERR gamma Coactivator Assay Kit, rabbit

1M DTT: Store at -20 °C or -80 °C

---

### Size

800 x 20 µL assays

---

### Kit Components

The TR-FRET ERR gamma Coactivator Assay Kit contains ERR gamma-LBD (GST) protein, fluorescently labeled PGC1α coactivator peptide, Tb-anti-GST antibody, and buffers. Store components as indicated in the assay protocol (-80°C, -20°C, or +4°C).

---

### Materials Required but Not Supplied

Microplate Reader

---

**Detection method** Fluorescent

---